thermionics laboratory, inc.

Remove screws



thermionics laboratory, inc.

Remove front plate



thermionics laboratory, inc.

Remove connector housing



thermionics laboratory, inc.

Remove screws



thermionics laboratory, inc.

Remove screws



thermionics laboratory, inc.

Remove pole piece



thermionics laboratory, inc.

Remove Coil



thermionics laboratory, inc.

Add a Shim between coil and front plate (optional)



Final Notes

- Some coils may be somewhat loose as they just sit in the e-Gun™. Use shims to take up space to dampen vibrations.
- After replacing coil, re-assemble in the reverse.

Thermionics Laboratory, Inc. 3118
Depot Rd.,
Hayward, CA 94545
Phone: 800-962-2310
Phone: (510) 538-3304
E-mail: sales@thermionics.com
www.thermionics.com